

Theory 5

Transition from Manual to automated visual Inspection



- Interpretation of inspection results and validation data
- Considerations on validation program for automated inspection
- Performance measurement
- Maintaining the manual inspection

Theory 5: Transition from Manual to automated visual Inspection

Some method comparison




Automated Visual Inspection (AVI)

- ✓ High speed and high capability
 - ✓ Highly reproducible
 - ✓ Consistent (no fatigue effect)
 - ✓ Defects presentation
-
- ✓ High initial investment
 - ✓ Works within strict condition (validated upstream process)
 - ✓ Detect “only” preset defects
 - ✓ Indiscriminative (i.e.: fiber and cracks are seen the same way)
 - ✓ Some uncovered area
 - ✓ Higher false reject rate



Semi-Automated Visual Inspection (SAVI)


- ✓ Adaptation
 - ✓ Speed
 - ✓ Brain
 - ✓ Flexible
 - ✓ Decision capable
-
- ✓ Inconsistent (fatigue effect)
 - ✓ Not highly reproducible
 - ✓ Susceptible to influence
 - ✓ Some uncovered area
 - ✓ Monotonous repeated work
 - ✓ Significant training effort



Manual Visual Inspection (MVI)

- ✓ Adaptation
 - ✓ Brain
 - ✓ Flexible
 - ✓ Decision capable
 - ✓ Classification of defects
-
- ✓ Inconsistent (fatigue effect, emotional)
 - ✓ Not highly reproducible
 - ✓ Susceptible to influence
 - ✓ Slow
 - ✓ Monotonous repeated work

Object presentation









Figure 1.1: The Perceptual Process



Chromatic continuity:
We see snow even when color changes drastically (RGB)

Theory 5: Transition from Manual to automated visual Inspection


Machine qualification : ICH Q9 - Risk base approach FMEA

AVI Process - AVI	Description	CQA / CPP																																													
 <p>Cameras inspection + for particles + for container defects</p> <p>Syringe with System closure, syringe with integer stopper, syringe without crack, syringe free of particle, syringe with correct</p> <p>Seidenader</p> <ul style="list-style-type: none"> - Syringe are transported by the conveyor to the Seidenader - Syringe are transported by several starwheels. Different cameras placed at different steps take pictures of different parts of the syringes. - The images are analyzed in order to detect any defective syringe based on the defined defects. - The syringes detected as defective are directed either to ejected syringes or rejected syringes. <p>Syringes detected as accepted move through a conveyor to the color code labeler.</p>	 <p>Description:</p> <p>Seidenader is the equipment used for automatic visual inspection in order to detect any SQIPP defect on syringe like particles, cracks, filling volume, closure system and stopper.</p> <p>The equipment detects and removes defective units with an acceptable rate and sustainable false-ejection/rejection rate.</p> <p>It is composed of :</p> <ul style="list-style-type: none"> - A transport system (frame, conveyors, wheels, tray etc.) - A vision system (lights, mirrors, cameras, SVIM module etc.) - A process control system (PLC, HMI, network architecture etc.) 	<p>For the Seidenader:</p> <ul style="list-style-type: none"> - CPP : See list below - CQA: Syringe with System closure, syringe with integer stopper, syringe without crack, syringe free of particle, syringe with correct volume, product's potency. <table border="1"> <thead> <tr> <th data-bbox="1091 456 1342 492">CPP</th> <th data-bbox="1342 456 1613 492">Control System</th> <th data-bbox="1613 456 1883 492">CQA</th> </tr> </thead> <tbody> <tr> <td data-bbox="1091 492 1342 549">Refeed transport mode</td> <td data-bbox="1342 492 1613 549">Stress free transport validation with a refeed rate</td> <td data-bbox="1613 492 1883 549">Syringe without crack</td> </tr> <tr> <td data-bbox="1091 549 1342 649">Synchronization Electrical phase = Mechanical phase</td> <td data-bbox="1342 549 1613 649">Alignment too</td> <td data-bbox="1613 549 1883 735">Syringe with System closure, Syringe with integer stopper, Syringe without crack, Syringe free of particle, Syringe with correct volume</td> </tr> <tr> <td data-bbox="1091 649 1342 735">Rotation profile</td> <td data-bbox="1342 649 1613 735">Global Document on high rotation specification</td> <td data-bbox="1613 735 1883 778">Product's potency</td> </tr> <tr> <td data-bbox="1091 735 1342 778">Rotation speed 4000 U/min</td> <td data-bbox="1342 735 1613 778"></td> <td data-bbox="1613 778 1883 821">Syringe without crack</td> </tr> <tr> <td data-bbox="1091 778 1342 821">Transportation</td> <td data-bbox="1342 778 1613 821">Stress free transport validation</td> <td data-bbox="1613 821 1883 1120">Syringe with System closure, Syringe with integer stopper, Syringe without crack, Syringe free of particle, Syringe with correct volume</td> </tr> <tr> <td data-bbox="1091 821 1342 863">Recipe tools and</td> <td data-bbox="1342 821 1613 863">Recipe check before production and PQ</td> <td data-bbox="1613 1120 1883 1163">Product's potency</td> </tr> <tr> <td data-bbox="1091 863 1342 906">Mirror xx position</td> <td data-bbox="1342 863 1613 906">Position control tool</td> <td data-bbox="1613 1163 1883 1206">Syringe free of particle</td> </tr> <tr> <td data-bbox="1091 906 1342 949">Camera positions</td> <td data-bbox="1342 906 1613 949">Luminance control tool Maintenance job description</td> <td data-bbox="1613 1206 1883 1249">Syringe without crack</td> </tr> <tr> <td data-bbox="1091 949 1342 992">Camera focus</td> <td data-bbox="1342 949 1613 992">Settings with access control</td> <td data-bbox="1613 1249 1883 1292"></td> </tr> <tr> <td data-bbox="1091 992 1342 1035">Luminance Intensity LED (Angle, Distance, Driver output parameter)</td> <td data-bbox="1342 992 1613 1035">SAP Control</td> <td data-bbox="1613 1292 1883 1335"></td> </tr> <tr> <td data-bbox="1091 1035 1342 1078">Access Control</td> <td data-bbox="1342 1035 1613 1078">SOP x</td> <td data-bbox="1613 1335 1883 1378"></td> </tr> <tr> <td data-bbox="1091 1078 1342 1120">Time out of refrigeration xx hours</td> <td data-bbox="1342 1078 1613 1120">Maintenance checklist</td> <td data-bbox="1613 1378 1883 1420"></td> </tr> <tr> <td data-bbox="1091 1120 1342 1163">Filter (Integrity, Presence, Cleanliness, Mounting)</td> <td data-bbox="1342 1120 1613 1163"></td> <td data-bbox="1613 1420 1883 1428"></td> </tr> <tr> <td data-bbox="1091 1163 1342 1206">Maintenance checking</td> <td data-bbox="1342 1163 1613 1206"></td> <td data-bbox="1613 1420 1883 1428"></td> </tr> </tbody> </table>	CPP	Control System	CQA	Refeed transport mode	Stress free transport validation with a refeed rate	Syringe without crack	Synchronization Electrical phase = Mechanical phase	Alignment too	Syringe with System closure, Syringe with integer stopper, Syringe without crack, Syringe free of particle, Syringe with correct volume	Rotation profile	Global Document on high rotation specification	Product's potency	Rotation speed 4000 U/min		Syringe without crack	Transportation	Stress free transport validation	Syringe with System closure, Syringe with integer stopper, Syringe without crack, Syringe free of particle, Syringe with correct volume	Recipe tools and	Recipe check before production and PQ	Product's potency	Mirror xx position	Position control tool	Syringe free of particle	Camera positions	Luminance control tool Maintenance job description	Syringe without crack	Camera focus	Settings with access control		Luminance Intensity LED (Angle, Distance, Driver output parameter)	SAP Control		Access Control	SOP x		Time out of refrigeration xx hours	Maintenance checklist		Filter (Integrity, Presence, Cleanliness, Mounting)			Maintenance checking		
CPP	Control System	CQA																																													
Refeed transport mode	Stress free transport validation with a refeed rate	Syringe without crack																																													
Synchronization Electrical phase = Mechanical phase	Alignment too	Syringe with System closure, Syringe with integer stopper, Syringe without crack, Syringe free of particle, Syringe with correct volume																																													
Rotation profile	Global Document on high rotation specification	Product's potency																																													
Rotation speed 4000 U/min		Syringe without crack																																													
Transportation	Stress free transport validation	Syringe with System closure, Syringe with integer stopper, Syringe without crack, Syringe free of particle, Syringe with correct volume																																													
Recipe tools and	Recipe check before production and PQ	Product's potency																																													
Mirror xx position	Position control tool	Syringe free of particle																																													
Camera positions	Luminance control tool Maintenance job description	Syringe without crack																																													
Camera focus	Settings with access control																																														
Luminance Intensity LED (Angle, Distance, Driver output parameter)	SAP Control																																														
Access Control	SOP x																																														
Time out of refrigeration xx hours	Maintenance checklist																																														
Filter (Integrity, Presence, Cleanliness, Mounting)																																															
Maintenance checking																																															

Theory 5: Transition from Manual to automated visual Inspection

Machine qualification : ICH Q9 – Example FMEA by block function

Potential Failure						Current Situation						Situation with appropriate measures			Situation after appropriate measure					Traceability				
2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21					
Process step	Potential Failure	Potential failure effect	Potential failure cause	Impact on CQA (xxx)	Current control measure	S	O	D	R	P	N	Appreciation	CPP Attribution xxx	Recommended preventive actions	Responsible	S	O	D	R	P	N	Appreciation	Reference / Parameter /SOP	
Material – Product																								
xx	Material	If the syringe has not the same structure -> recipe will not analyze correctly -> High false ejection (example: flange variability)	Variability on the material design	No	Supplier notification management (Change control) and yearly business review	2	3	3	18	⚠️	N/A	N/A	N/A	2	3	3	18	⚠️	N/A					N/A
xx	Product	Change behavior of mobile particles or air bubbles -> missed particles.	Product viscosity do not fit the specification	S4	AQL	3	1	3	9	✅	Rotation profile	N/A	N/A	3	1	3	9	✅	N/A					N/A
xx	Product	Luminance and rotation impact are too high -> Illumination energy and Shear stress destruct components inside -> Strength of product decreased.	Product stability do not fit the specification	S6	Recipe detection Quality control Fixed parameter in the recipe -> List of Global Document on High Rotation Specification is given for each product.	5	1	1	5	✅	Rotation profile and luminance intensity	PE done x for xx product, machine and parameter	N/A	5	1	1	5	✅	N/A					N/A
xx	Product	Product not well homogeneous -> False high ejection rate	Sedimentation of the product does not fit the specification (offline production)	no	Tub is slightly turned to let the product been homogenized by operator. Prerotation step in Seidenader before CSI Station	2	2	3	12	⚠️	N/A	Define the process for offline production in SOP xxxx	N/A	2	1	1	2	✅	N/A					N/A



Installation Qualification

- Documentation verification , component data verification, drawings, system Installation verification , utilities, Software and IT verification

Operational Qualification

- HMI Layout verification
- Alarms verification
- Screen navigation, access verification, security verification
- ER/ES verification
- MES (Manufacturing Execution System) server communication
- Backup / Restore and disaster recovery
- Containers handling
- Counters and cells control
- VI rotation at 360° control
- Recipes version verification

Since the particulate visibility statement in the XIX Revision of the Pharmacopeia (9) is based upon a deterministic human inspection it is inappropriate and should be discarded. With both manual and automated systems regarded as probabilistic, they can now be similarly evaluated and their demonstrated capability rigorously compared.

longevity estimates. The particular containers rejected in any single inspection cannot be accurately predicted except for two special cases: those containers that are absolutely clean and are never rejected and those containers with gross defects that are rejected in every inspection.

In terms of the two-dimensional probabilistic inspection model, Uhlir utilized two unrelated one-dimensional probability distributions: manual and machine. In consequence, the differing sensitivities of the two methods can yield the Venn diagram result shown in Figure 1. Here, the manual inspection and the automated device perform in exemplary fashion. Figure 1 indicates, however, that the sets of containers rejected by each method had few containers in common. This comparison suggests that the Uhlir evaluation methodology may not generate the demonstration of equivalence that CGMP's require in the validation of alternative inspection methodologies and devices.





Figure 1—Venn diagram of two inspector particulate inspection demonstrating the expected paradoxical results.

Theory 5: Transition from Manual to automated visual Inspection

Why Correlogram unit by units does not make sense ?



Comment R Veillon

When J Knapp draw a correlogram of between 2 method, each plot is the number of units in each probability class

That is NOT paired comparison per DEFECT

The capability of one process relative to the other cannot be evaluated until the correlation between the results of both inspections is established. This correlation is based on an examination of the inspection history of each container in each inspection process. Sufficient inspection replications are required to assure statistically reproducible results with acceptable tolerance intervals. Since we are dealing with probabilistically defined quantities, statistical tools must be used. The basic questions of replicability, relative per-

Figure 3—Correlogram comparing the results of 72 manual and 70 PDA/A inspections. A comparison summary of the two inspection methods is included in the computer printout. Of major interest is the fact that only 1 (MRZR1) of the 52 were rejected manually with a probability of 1.0. The PDA/A rejected 20 (ARZR1) of the 52 with a probability of 1.0. The plusses on each axis are the 10% increment points from a rejection probability of 0 to 1.0. The abscissa is for the automated system; the manual system rejection probability is the ordinate. The dotted lines shown are the Reject Zone boundaries for both systems. The * symbols indicate a single container at a point in the plane, a Y indicates a number of containers greater than 9. Values between 2 and 9 are shown directly. The reject rate, R, the Reject Zone Efficiency, RZE, and the undesired reject rate in the Accept and Gray Zones, RAG, are tabulated under the histogram with suffix 1, 2, and 3 to indicate sequential inspection number. The prefix N indicates manual inspection; the A prefix indicates an automatic inspection.

Theory 5: Transition from Manual to automated visual Inspection

Classification of defects by « probability sub group »





Figure A2—One-dimensional histogram of a normal batch showing the accept Gray and Reject Zones defined by the human based standard inspection.

Figure 2—Histogram of probability of detection for a 1000 vial randomly selected test group. The Schering standard 17 second paced manual inspection was employed.

To accomplish this evaluation, two random distributions must be compared.

When the implications of the two dimensional probability plane of Figure A2 are examined it becomes apparent that each entry in either system can be transformed into a distribution in the other system.

TABLE AII. Probabalistic Distribution of Rejection Probabilities for Containers in "Manual" Inspection and "System" I and II

	N(0)	N(.1)	N(.2)	N(.3)	N(.4)	N(.5)	N(.6)	N(.7)	N(.8)	N(.9)	(N1.0)
'MANUAL'	1	1	1	1	1	1	1	1	1	1	1
'SYSTEM'	2.5	1.5	1.5	.5	0	0	0	.5	.5	1.5	2.5
I & II											


$$RZE(M1) = \frac{RZR(M1)}{RZN} = \frac{14.7}{18} = 81.7\% \quad (\text{Eq. 4})$$

RZE(Mn) = efficiency of rejection in Reject Zone

RZN(Mn) = number of vials identified in the manual Reject Zone


RZR(Mn) = Reject Zone reject quantity as defined in manual inspection

$$RZE(M2) = \frac{RZR(M2)}{RZN} = \frac{12.2}{18} = 67.7\%$$



Material and Methods











- [10 kits + good units] = 1 inspection lot order
- No information given to inspectors = routine inspection
- No interactions with inspectors to avoid any interferences
- Changed shift to avoid interactions between inspectors
- 1 inspection every day during 2/3 weeks, one inspector at a time
- Kit verified every day for defect state, replaced broken units to identical
- QF Result compiled for each inspector



3. Material and Methods

- Data reporting

QF = number of ejected / number of inspected

	Operators	 1	 2	 3	 4	 5	 6	 7	 8	 9	 10
KIT	DEFECT										
Kit A	Defect #1										
Kit A	Defect #...										
Kit A	Defect #nn										
.....										
Kit J	Defect #1										
Kit J	Defect #...										
Kit J	Defect #nn										

QF #1A

.....

QF #nn


QF #1J

.....

QF #nn

MRZE¹⁵

QF distribution for Syr. Particle defects





Data from particle MRZE studies
2011+2014 WN Syr.

Key learning:


Particle detection in MVI is highly probabilistic: operator variability is lower with very high QF > 0.70
Operator variability higher with lower QF [0.3:0.8]

AVI qualification by Knapp



=> Comparison of 2 distributions of number of unit having same detection rate

ADR vs. QF



Manual Detection rate

Key learning:

Particle detection in AVI has a higher ADR and is less probabilistic than MVI

Specially in range of QF > 0.70

In range with Lower QF

ADR is higher than MVI but more heterogeneity between particles (floating/precipitating)



Theory 5: Transition from Manual to automated visual Inspection

Why is it important to maintain MVI ?

- AQL done in MVI
- AVI qualification is compared to MVI reference

Key take away:

- In this section you have learnt:

AVI

Machine qualification

VS

Interpretation of inspection results and validation data : Knapp review

MVI

Considerations on validation program for automated inspection

Performance measurement

Maintaining the manual inspection
