



Instructor Lead: Romain Veillon Sr Manager GSK

Vaccines:

contact romain.veillon@gsk.com

- **Theory 1: Introduction to regulatory landscape of visual inspection**

- USP 1, USP 788 and 1788, USP 790 and 1790
- PhEur e.g. 2.9.20
- JP e.g. 6.06
- Annex 1
- Similarities and differences in compendial methods
- 100% inspection and AQL testing
- Definitions and practical examples of inherent, intrinsic and extrinsic particles
- Examples of regulatory citations 483s




Mastering Automated Visual Inspection


Theory 1: Introduction to regulatory landscape of visual inspection



Technology




Quality / regulatory



Business



- 10 Golden rules for VI.....



- USP<1>
- USP<788> particle definition
- USP<790>and <1790>
- PhEur e.g. 2.9.20 vis
- JP e.g. 6.06
- Annex 1: new draft for comment dec 17
- Similarities and differences in compendial methods
- 100% inspection and AQL testing
- Definitions and practical examples of inherent, intrinsic and extrinsic particles





- USP<1790> Effective Aug. 2017

- New Annex 1 draft for comment dec 2017



Adobe Acrobat
Document



Adobe Acrobat
Document



2019 PDA Visual Inspection Forum

Preparing for the Future of Visual Inspection



CONNECTING
PEOPLE
SCIENCE AND
REGULATION™

APRIL 23-24 | WASHINGTON, DC

EXHIBITION: APRIL 23-24

AN INTRODUCTION TO VISUAL INSPECTION TRAINING COURSE: APRIL 25-26

pda.org/2019Visual

#PDAVisual

Inspection trends and market recall for Visual Inspection

Romain Veillon

Sr Manager Visual Inspection MSAT

GSK Vaccines




Disclaimer

***The content, material, perspectives and opinions expressed in this presentation are solely those of the presenter, and do not reflect the views or opinions of PDA.
Conflict of interest: Romain Veillon is an employee of the GSK group of companies.
This work was sponsored by GlaxoSmithKline Biologicals SA.***



Inspection trends with Recent FD 483s



Note: the findings statements reported after are excerpts of a list released under the Freedom of Information Act and published by the FDA on FDA website. Those findings are anonymized.

Defect kits

2009

« There is **no standardized kit of defects for the training of employees who conduct visual inspection of the vials components** “

2013


“Your operator's visual inspection of Lyophilized drug product qualification program **does not include examples of glass particulate in vials for training purposes.**”

2014

“The defect sets utilized for qualification of inspection **do not have defects that are representative of defects found in routine inspection, retention examination, and complaints**”

2018

“There are no documented procedures or reference for the creation of the visual inspectors qualification kits. There are no established specifications for the size of the particle included with the kits”



USP<1790>

- 7.1 Standards
- 7.2 Preparing Defect Standards
- 7.3 Particle Types

Test sets

2016


« Qualification of visual inspectors and validation and verification of the ...inspection system are not based on well-characterized test sets.

No written procedure has been established to ensure test sets for visual inspection include particles in the visible size range similar to production rejects other than amicron glass particle.

No record documenting the creation of test sets used for qualification of visual inspectors was provided....”

2017

“Defects that typically occur during production are not characterized in sufficient detail to allow for consistent creation or selection of defects to include in test sets used for qualification of inspectors. Additionally, during creation of defect test sets, defects in the test sets are not well-characterized to ensure they are representative of typical production defects. For example, there is no information on particle size, particle material type (for light and dark particles), crack size, and crack location.”



USP<1790>

- 7.1 Standards
- 7.2 Preparing Defect Standards
- 7.3 Particle Types


Multiple stage inspection

2015

« Lots of finished drug product that fail the initial automated visual inspection limit on the system can then be re-inspected using the semi-automated manual system .There are no establish limits for the re-inspection of lots of product that fail the initial inspection that are then re-inspected...”

2018

“You have not adequately assessed spinning parameters, such as rotation per minute (RPMs) of your semi-automated inspection equipment which affect the capability of your visual inspection process.”




USP<1790>

- 3. TYPICAL INSPECTION PROCESS FLOW
- 3.3 Remediation and Alternative Practices- two stage inspection
- 6.2 Semi-Automated Visual Inspection

MVI Baseline performance studies

2017

“The probability of detection of particulates used in the defect test sets for manual visual inspection has not been determined to qualify these defects for use in the test set.”



USP<1790>

- 7.4 Rejection Probability Determination
- 7.6 Types of Test Sets – Threshold studies

Standard work MVI


2011

« The visual inspection certification program is not adequate :, no rotation of the unit required during the visual inspection..”

2016

“During the visual examination, I observed that the operator's visual inspection process was inconsistent in the amount of inspection time she spent on drug product units.”

“SOP ... does not instruct the visual inspector to gently swirl and invert the container during visual inspection”



USP<1790>

- 6.1 Manual Visual Inspection – critical process parameters in MVI

Certification conditions

2011


“procedures and current practices for the certification of the operators conducting visual inspection were found not representative of current production conditions »

2012

« ..The qualification does not entail reviewing and identifying defects under the same conditions that during manufacturing operations “

2016

“The certification exercise does not simulate conditions observed during routine product inspection operationsTherefore, the current certification/re-certification procedure does not challenge the capability of the operators to recognize and separate all types of defective vials during the maximum individual inspection interval ”



USP<1790>

- 7.7 Training and qualification of human inspectors

Certification conditions


2017

“Routine visual inspection occurs on-line with operators
..... During the qualification the operators work offline ...”

2018

“however these photographs are unclear and inadequate to identify glass particles in vials. Without an adequate qualification vial, your firm cannot ensure your operators can observe this defect during 100% visual inspection.

.... your firm does not have a procedure to address an employee who repeatedly failed to identify a specific defects during all operator qualification runs”



USP<1790>

- 7.7 Training and qualification of human inspectors


AVI Validation

2014

“Data was not available at the time of the inspection to demonstrate that the has been qualified as equal to or better than the inspection...”

2017

« There is no documentation of Process Qualification study of the ... machines capabilities to detect vials heel crack defects. ”



USP<1790>

9. Conclusion –

.....Where machine methods are used, the equipment must be validated to demonstrate equivalent or better performance when compared to manual inspection.

6.3 Automated Visual Inspection

.....Significant effort is required to program these systems and to test their performance against a range of known defects,

Machine Maintenance / Daily test

2014

“Vials could be heard hitting against each other during the addition... This practice potentially subjects post visually inspected vials to damage. “

2015


« The light intensity of each unit is not verified during routine preventive maintenance and is not verified prior to use.

The functionality test used to determine the reject function of the equipment is required before and after 100% visual inspection.

The functionality test results for each equipment is not clearly documented as to the test results. ”

2015

“The equipment are used to perform 100% visual inspection of lyophilized vials, For example,.... **The light intensity of each unit is not verified during routine preventive maintenance and is not verified prior to use.** “



AQL QUALITY OVERSIGHT

2011


“There is no direct QA oversight of the operators performing the visual inspections. Operators are only observed during their certification process, but not on a routine basis.”

2015

« Quality oversight over visual inspection is deficient. For example,

a. AQL inspections are conducted by personnel that also perform the 100% visual inspection

b. From September 2013 to September 2015, QA oversight over the 100% visual inspection operations has occurred six times.”



Action after AQL failed

2015

« There was no tightened 100% inspection performed for this lot even though the initial AQL failed for a Major defect. “

2015


“There are no established limits for the number of times any single lot can be re-inspected. Additionally, there are no tightened limits established for the re-inspection”

2016

“reported a particle identified in a vial during an AQL inspection. There was no documentation on the identity of the particle and whether it was inherent or foreign (black debris, fiber, glass fragments, etc.).”

2015

“there is no requirement to tighten the inspection limits or increase the sample size for the second AQL inspection”



Trending

2011

« There is no written SOP that include performing trend evaluation to determine the root cause that created the quality related attributes”

2013


“There is no tracking or trending of the number of xx vials initially rejected as “Particulate Fiber” ”

2016

“inspection system, does not have an overall alert/action limit for total rejections”

2018

“You do not monitor long term drift during your establishment / re-establishment of in-process limits.”




Continuous improvement

2015

« According to SOP, the test set library shall be largely covered with regards to existing (i.e., known) defects. No less than eight deviations for cracks on vial bottoms occurred since approximately, for example, deviation, **This defect type has not been added to the test set library to date**”

2015


“**Particles size was not determined to facilitate assessment of the reliability of detection during visual inspection**”



Market recall trends for Visual Inspection

VI Recall trends for US Market


US FDA Recall Notices (All)



Prepared by John Shabushnig, Insight Pharma Consulting from Recall Archive data on fda.gov

VI Recall trends for US Market


US FDA Drug Recall Notices



Prepared by John Shabushnig, Insight Pharma Consulting from Recall Archive data on fda.gov

VI Recall trends for US Market

US FDA Parenteral Drug Recall Notices



Prepared by John Shabushnig, Insight Pharma Consulting from Recall Archive data on fda.gov

VI Recall trends for US Market


Visible Particulate Recall Notices



Prepared by John Shabushnig, Insight Pharma Consulting from Recall Archive data on fda.gov

VI Recall trends for Canadian Market Focus on Parenteral for 2017

HEALTH CANADA 2017



Adverse event	1
CCI	4
mixup	1
Packaging	3
QC testing	5
Sterility	13
storage	3
VI	2
Grand Total	32


All VI recall are linked with
Particulate Matter issues

VI Recall trends for Japan Market

Focus on Parenteral from 2012 to 2018

Parenteral recalls are 10% of all Drug recalls

Japan Market Parenteral Recall 2012-2018 YTD



Category	2012 Jan-Dec	2013 Jan-Dec	2014 Jan-Dec	2015 Jan-May	2016 Apr-Dec	2017 Jan-Dec	2018 Jan- Sept 20	TOTAL
Particle*	1	0	3	1	0	0	2	7
Primary packaging**	0	0	2	0	0	1	1	4
OoS	3	0	1	0	6	5	5	18
Non-compliance	0	9	4	1	3	4	4	25
Incorrect labeling	0	3	2	0	1	1	0	7
Grand Total	4	12	12	2	10	11	10	61
Number of Recall of all Drugs	81	135	82	25	77	100	87	587


Prepared by Katayama San based on PMDA (Pharmaceuticals and Medical Devices Agency) of Japan, lacking some data in 2015-16

VI Recall trends for Japan Market


Focus on Parenteral from 2012 to 2018



PARENTERAL RECALLS JAPAN 2012-2018 YTD





PARENTERAL RECALLS JAPAN PARTICLE 2012-2018 YTD



Prepared by Katayama San based on PMDA (Pharmaceuticals and Medical Devices Agency) of Japan, lacking some data in 2015-16

VI Recall trends for French Market Focus on 2017 for Parenteral




Prepared by Romain Veillon, from Recall Archive data on ANSM website

VI Recall trends for Swiss Market Focus on Parenteral recall 2016-2018

CCI	4
packaging	4
process issue	3
QC testing	2
VI	2
Grand Total	15

PARENTERAL RECALL SWISS MEDIC 2016-2018 YTD



Thanks you for your attention

Many thanks to contributors :

- **John Shabushnig** for FDA recall analysis
- **Hirohito Katayama** for Japan recall analysis

Sources:

FDA (USA): <https://www.fda.gov/Safety/Recalls/ArchiveRecalls/default.htm>

Swiss Medic

(Switzerland): <https://www.swissmedic.ch/swissmedic/fr/home/humanarzneimittel/marktueberwachung/qualitaetsmaengel-und-chargenrueckrufe/chargenrueckrufe.html>

ANSM (France): <https://www.ansm.sante.fr/S-informer/Informations-de-securite-Reclams-lots-et-de-produits>

Health Canada: <http://www.healthycanadians.gc.ca/recall-alert-rappel-avis>

Japan: Sourced from H.Katayama

Contact: romain.veillon@gsk.com





- Thanks you for your attention
- Contact: romain.veillon@gsk.com

Many thanks to contributors :

- John Shabushnig for FDA recall analysis
- Hirohito Katayama for Japan recall analysis
- Justine Donfack for 483s analysis during her traineeship