

# Understanding Sterilization

- Sterilization basics
- Radiation Technology & Ethylene Oxide



**ANNICK GILLET**  
TECHNICAL DIRECTOR, EO PHARMA




- **Basics of sterilization**
  - Distinguish disinfection, sterilization and decontamination
  - Definition
  - Selection of sterilization method
  - Difference between Aseptic Assembly and Terminal Sterilization
  
- **Sterilization using Irradiation**
  - Gamma
  - E-Beam
  
- **Sterilization by Ethylene Oxide**
- **Comparison between technologies**

# Sterilization Basics


- Decontamination Vs Sterilization
- Terminal Sterilization Vs Aseptic Assembly
- Method selection

## Decontamination Vs Sterilization



<b>Validation</b>			
<b>Sterilization</b>	<b>Decontamination</b>	<b>Disinfection</b>	<b>Cleaning</b>
<p>The application of a lethal sterilizing agent to finished product within a sealed container to achieve a predetermined <b>sterility assurance level (SAL)</b> of <math>10^{-6}</math> or better –</p> <p><i>GMP Annex 1 Draft</i></p>	<p>A process that <b>eliminates viable bioburden</b> via use of chemical agents</p> <p><i>GMP Annex 1 Draft</i></p>	<p>The process by which <b>surface</b> bioburden is <b>reduced</b> to a safe level</p> <p><i>GMP Annex 1 Draft</i></p>	<p>Removal of <b>contamination</b> from an item to the extent necessary for further processing or for intended use</p> <p><i>ISO 11139:2006</i></p>

**A sterile product is one that is free of viable microorganisms**



**Absolute sterility can never be guaranteed !**

- 100% control of the batch is not possible
- No assurance that any microorganism can be detected during Sterility Test

**Sterility Assurance Level (SAL)** = The **probability** of a single item in a batch being non-sterile after being subjected to a sterilization process.

**Sterile: SAL ≤ 10<sup>-6</sup>**

**SAL likelihood of surviving organisms**

10<sup>-1</sup> = 1:10


10<sup>-2</sup> = 1:100

10<sup>-3</sup> = 1:1,000

10<sup>-4</sup> = 1:10,000

10<sup>-5</sup> = 1:100,000

**10<sup>-6</sup> = 1:1,000,000**



## Sterility is much more than just a process!

### Initial contamination level

- Microbiological status raw material and components
- Cleaning and disinfection procedures
- Environment control
- Personnel Hygiene



### Equipment

- Control
- Maintenance
- Calibration




### Product preservation

- Packaging
- Storage





## Pharmaceutical Product Life Cycle




**Think about sterilisation as soon as possible during product development**

**Sterile means** : Safe Product & Functional product




**Selection of the right sterilization method is critical !**

**There are two (2) methods to produce a sterile drug product:**




## Aseptic Assembly



**Maintain** sterility of a product that is assembled from components, each of which has been previously sterilized

# Sterile


## Terminal Sterilization



**Exposure** to a physical or chemical sterilizing agent for a predetermined extent of treatment

# Sterilized

## Is the effectiveness of a sterilization process assessed the same way for AA or TS products?



*Reference: ISO TS  
19930:2017*

## **Probability of Non Sterile Unit (PNSU):**

The probability of one or more microorganism being present on a product item in a population of items.

- No SAL can be calculated for aseptically assembled products (no killing kinetic)
- Rely on multiple factors
- Estimated by Media Fill Test (process simulation)
- Bracketing approach (MFT before and after process)




## Selection of the Sterilization Method:



**“Wherever possible, a process in which the product is sterilized in its final container (terminal sterilization) is chosen”**

European Pharmacopoeia 9.7

*Per PDA 2017 Survey – 30% of Aseptically assembled product could be Terminally sterilized !*



## Selection of the Sterilization Method:

Use a **structured approach** to select the most appropriate sterilisation method

Based on EMA - CPMP/QWP/054/98 Decision  
*Tree for the selection of sterilisation methods*




Prior to making your choice, consider mitigation options:

- Can your **formula** be adapted ( limit degradation and impurities)?
- Can the **container** be adapted ?
- Can you select compatible **component** with selected sterilization process ?
- Can the **process** can be optimized (limit impact)?




*CPMP/QWP/054/98 Decision Tree for the selection of sterilisation methods*



## Ethylene Oxide

- (EO) gas



## Irradiation

- Gamma ray
- Accelerated electrons (E-beam)
- X-rays



## Other

- Moist heat
- Dry heat
- Vaporized hydrogen peroxide (VHP)
- Gas plasma
- Low temp Steam Formaldehyde (LTSF)
- Nitrogen Dioxide (NO<sub>2</sub>)

### Most common methods

for terminal sterilization of single use medical devices


# Radiation Technology

- General principles
- Gamma
- E-Beam
- Sterilization validation

## General Terminology

### Radioactivity:

Electromagnetic radiation (photons) produced by radioactive decay.



**E-beam** = Electrons (with a mass)

## General Terminology

### Radiation

Energy in the form of waves or moving subatomic particles

### Radioactive

Substance emitting radiation

### Irradiation


Exposure to radiation  
≠ Making something radioactive



## General Terminology

### Ionising Radiation

Radiation capable of knocking electrons out of their thermal orbits in atoms or molecules



### (Absorbed) Dose

Measure of the amount of energy that is absorbed by the material while exposed to a radiation source.

**Unit: Gray      1 Gy = 1 Joule per Kg material**

## Irradiation process monitoring:

### Dosimeter

Device having a reproducible, measurable response to radiation, which can be used to measure the absorbed dose in a given system.



0 kGy

12 kGy


25 kGy

50 kGy

0kGy

# Sterilization by Irradiation – General principles

Two methods to generate irradiation :





## Gamma Irradiation



# Sterilization by Irradiation : Gamma

Scale of irradiation :



**Sterilization Dose  
10.000 – 40.000 Gy**

Dose that may cause symptoms of radiation sickness (1000 mGy)



1000

500


100

50

10

1

Typical chest X-Ray (0,1 mGy)




Effects of ionizing Radiation on DNA

**Source:**  $^{60}\text{Co}$  (mostly)

**Decay rate:** 12% per year (Half life 5,3 years)

**Source Activity:** Several Million Ci




Isotropic radiation flux



## Source Rack


Cobalt-slugs in a source pencil




Source module




Source rack



## Layout Gamma facility



## E- Beam irradiation





## Electron Beam

Directed stream of electrons (B radiation) produced by a particle accelerator

## Beam energy


Speed of the electrons. Parameter related to depth of penetration

Limited to 10 MeV for medical device sterilisation (ISO 11137-1) to avoid radioactivity induced in product




IBA Rhodotron

## Layout E-Beam facility





## Electron Beam & Gamma, Penetration



Parameter	Gamma	E-Beam
<b>Irradiation parameter</b>	Cycle Time Density	Conveyor speed Density Scan width Beam energy
<b>Radiation Field</b>	Isotropic	Highly directional
<b>Geometry of material and heterogeneity of Product</b>	Important to consider	Critical

# Sterilization by Irradiation : comparison

Parameter	Gamma	E-Beam
<b>Product Treatment</b>	Pallet/Tote	Boxes
<b>Dose Rate (Dmin 25KGy)</b>	Hours	Seconds
<b>Dose uniformity ration (DUR)</b>	Low sensitivity to product thickness	sensitivite to product thickness
<b>On/Off Technology</b>	No	Yes
<b>Flexible Target Dose</b>	No	Yes
<b>Process validation</b>	Straightforward	Potentially complicated

## Relevant Standards:

### ISO 11137-1:2015

Sterilization of health care products – Radiation – Part 1: Requirements for development, validation, and routine control of a sterilization process for medical devices


### ISO 11137-2: 2015

Sterilization of health care products – Radiation – Part 2: Establishing the sterilization dose


### GMP – Annex 12

Use of ionising radiation in the manufacture of medicinal products


## Method $VD_{max}$



Standard Distribution of resistances (SDR)



# Sterilization by Irradiation: validation principles



# Sterilization by Irradiation: validation principles


Bioburden is critical parameter in Irradiation technology

Sample Item Portion (SIP) is frequently used for bioburden evaluation .

Basis for SIP can be:



### Length

- Consistent diameter tubing




### Mass

- Powders
- Gowns
- Absorbable implants




### Volume

- F



### Surface Area

- Non-absorbable implants
- Variable diameter tubing



## Select Sterilization Dose

Method  $VD_{max}$

Example minimum  
Dose to apply related  
to bioburden

Bioburden Range	Dose (kGy)
$\leq 0.1$ to 1.5	15.0
$\leq 0.1$ to 9.0	17.5
$\leq 0.1$ to 45	20.0
$\leq 0.1$ to 220	22.5
$\leq 0.1$ to 1000	25.0
$\leq 1.0$ to 5000	27.5
$\leq 1.0$ to 23,000	30.0
$\leq 1.0$ to 100,000	32.5
$\leq 1.0$ to 440,000	35.0



- Select Verification Dose:  $VD_{max}^{25}$



Bioburden	Verification Dose (kGy)
40	8.6
<b>45</b>	<b>8.7</b>
50	8.8
55	8.9


Verification is conducted at an SAL of  $10^{-1}$  with 10 product items irradiated.

## Dose Mapping

Establish the distribution of absorbed dose within the irradiation container when packed with product in a defined configuration

- Min and Max limits of absorbed Dose
- Define cycle time
- Establish monitoring points

-  Min Dose = 28K Gy
-  Max Dose = 37K Gy



## Quarterly Dose Audit (QDA)

### Check bioburden

Can vary due to

- Season
- Environment ...



### Verification Dose

(Often In a research irradiator)

Ex: 8,7KGy



### Sterility Test

SAL $10^{-1}$



### Confirm Product




SAL  $10^{-6}$

With Routine Dose

Ex:25KGy




# Sterilization by Irradiation: examples



## ... But also



Grafts



API

## Summary


**Minium & Maximum dose to product shall be defined**

**Methods 1, 2, VDmax, “equivalent method”**

**Based on natural product bioburden**

**Routine process monitored with dosimeters**


**Quarterly Dose Audit (QDA) required**



# Ethylene Oxide Sterilization

## Introduction

# Sterilization by Ethylene Oxide : History



Ethylene Oxide discovered

Charles Wurz

1859



First production of Ethylene Oxide

Union Carbide Chemicals

1925



Patent for sterilization of spices

Lloyd Hall

1938



Use in sterilization of materials

1940




Dr. Lloyd Augustus Hall, a food scientist, while working for Griffith Laboratories, devised a process known as the Ethylene Oxide Vacugas treatment to control the growth of molds and bacteria. Griffith and Hall received US Patent 2,189,949 in 1940.



## Properties


- Toxic gas
- “Sweet smell” from ca. 500 ppm concentration
- Forms with air explosive mixtures (2.6 %)
- Oncogenic by inhalation
- Irritating for skin and respiratory system
- Mutagenic for animals and very likely for humans



Last choice but sometimes the only one !

## Mode of Action

- Extremely reactive
- Irreversible reaction with DNA and proteins (alkylation)
  - The molecule loses function
  - Replication stops
  - The cell dies



## Mainly used to sterilize:

- Heat-sensitive material
- Material sensitive to ionizing radiation
- High Volumes
- Packs with multiple components




## Device/packaging must be permeable to the gas

- No aqueous substances
- No protein-type materials
- Powders, batteries, electronic circuits have to be assessed (risk of explosion)
- Vacuum/heat can have adverse impact on some packaging (bubble wrap packaging, polystyrene)



# Sterilization by Ethylene Oxide

## Critical Parameters for Effective EO Treatment




## Temperature (T)

**EtO kills microorganisms** even at temperatures below 10°C (50°F)



**Industrial sterilization** performed in 40-60 °C (104–140°F) temperature range

**Q10 Effect** increase by 10°C (18°F) = 2x Deactivation Rate



**Temperature increase** may increase permeability of gases through materials



## Relative Humidity (RH)



## EO-Concentration



**At constant T and RH – if EO concentration increases microbiological Deactivation is more effective - up to c. 800 mg/l**


- ~ 500 mg/L @ 131°F
- ~ 800 mg/L @ 86°F



## Time

**Microbiological deactivation**  
is more effective  
with longer gas  
dwell phase

**Industry cycles**  
2 to 10 hours gas  
dwell phase  
Typically 3-4 hours



## Customer Needs To Define


Product  
Families/Processing  
Categories

Finalize Packaging

Load Configuration

Bioburden

Internal PCD



## Sterilization Process




3 key phases




## Preconditioning


- 40-45°C
- 45–75 %RH
- 10-24 hours



## Typical EO Cycle Design – Deep Vacuum




**GENERIC CYCLE**




# Sterilization by Ethylene Oxide



Aeration




- 45°C
- Forced circulation
- 1-3 days



# Sterilization by Ethylene Oxide

## Monitoring EO Sterilization - Biological Indicators

- Usually, the BI contains at least a million spores (>10Exp6) of an organism that is highly-resistant to the EO process (*Bacillus atrophaeus*)
- Growth is very characteristic (orange ring)




**Spore**



## Process Challenge Device (PCD)

**Item designed to constitute a defined resistance to the sterilization process and used to assess performance of the process**


- Internal PCD (IPCD)
- External PCD (EPCD)



# Sterilization by Ethylene Oxide

## Monitoring EO Sterilization – Biological Indicators

We design the validation to show that the **BI** is more difficult to kill than natural occurring bioburden




## D Value

The Time needed to deactivate 90% of population of microorganisms (or 1 Log Reduction)

$$\text{SAL} \leq 10^{-6}$$

*The sterilization cycle is **validated** to predict achievement of an SAL equal to or less than a specified value ( $\geq 12\text{LR}$ )*




## Level of Sterility Assurance

*Example:*

$$D_{value} \text{ IPCD} = 15 \text{ min} = 1 \text{ LR}$$

6 LR = 90 min (Half cycle)

12 LR = 180 min (Full cycle)




# Sterilization by ETO: validation principles




- Establish Product/IPCD  $D_{Value}$
- Product Natural bioburden killed
- Define Challenges (IPCD -EPCD)



- Confirm IPCD selection ( $SAL \leq 10^{-1}$ )
- Confirm External Challenge (EPCD)




- $SAL \leq 10^{-6}$
- Aeration validation - Residue Tests



## Compounds that remain on product after EO sterilization


- Ethylene Oxide (EO)
- Ethylene Chlorohydrin (ECH) = EO + HCL
- Ethylene Glycol (EG) = EO + H<sub>2</sub>O



**ISO 10993-7:2008**

**“Biological Evaluation Of Medical Devices-  
Part 7: Ethylene Oxide Sterilization Residuals”**

There are Three Patient Exposure Categories



## ISO 11135:2014

Sterilization of medical devices – Requirements for the development; validation and routine Control of a Sterilization Process for Medical Devices – Ethylene Oxide

## ETO Residuals

*ISO 10993-7:2008 (R) 2012*

Biological evaluation of medical devices - Part 7: Ethylene oxide sterilization residuals

## GMP – Annex 1 (Draft)

- Manufacture of Sterile medicinal Products

## Bioburden

*ISO 11737-1:2018*

Sterilization of medical devices (Microbiological methods) Part 1: Determination of a population of microorganisms on products



## Product Sterility

- *ISO 11737-2:2009 (R) 2014*

Sterilization of medical devices (Microbiological methods) Part 2: Tests of sterility performed in the definition, validation and maintenance of a sterilization process

- *United States Pharmacopeia (USP) Chapter <71> Sterility Tests*
- *European Pharmacopeia (EP) Chapter 2.6.1 Sterility*
- *Japanese Pharmacopeia (JP) Chapter 54. Sterility Test*

## Biological Indicator Tests

- *ISO 11138-1:2017*

Sterilization of health care products (Biological indicators) Part 1: General requirements

- *ISO 11138-2:2017*

Sterilization of health care products (Biological indicators) Part 2: Biological indicators for ethylene oxide sterilization processes

- *ISO 14161: 2009 (R) 2014*

Biological indicators. Guidance for the selection, use and interpretation of results

## Medical Devices



Surgery packs



Catheters



vials



Bandages


## Drug products



API



Prefilled syringes  
(external)



Active Implants



Auto-Injector  
(external)

# Sterilization : Comparison Radiation & Ethylene Oxide

Parameter	Gamma	E-Beam	EO
Process	Individual product, box, tote, pallet	Boxes	Pallets
Material compatibility	Not compatible with some type of polymers (PTFE and PVC affected)	Wider polymer compatibility compared to Gamma	Very good No liquid/proteins Low Temperature (40-55°C)
Validation	Straightforward	Straightforward	Complicate
Validation principle	Based on bioburden	Based on bioburden	Based on Bio Indicators
Requalification	Every 3 months (QDA)	Every 3 months (QDA)	Every 2 years (1 cycle)
SAL	<10exp6	<10exp6	<10exp6
Residues	None	None	ETO,ECH,(EG)

# Sterilization : Comparison Radiation & Ethylene Oxide

Parameter	Gamma	E-Beam	EO
Process	Individual product, box, tote, pallet	Boxes	Pallets
Material compatibility	Not compatible with some type of polymers (PTFE and PVC affected)	Wider polymer compatibility compared to Gamma	Very good No liquid/proteins Low Temperature (40-55°C)
Tolerance for density variation	High	Low	Medium
Routine monitoring	<ul style="list-style-type: none"> <li>• Only a few parameters (Time, Size, density)</li> <li>• Dosimeter</li> </ul>	<ul style="list-style-type: none"> <li>• Higher Nb of parameters</li> <li>• Dosimeter</li> </ul>	<ul style="list-style-type: none"> <li>• Multiple cycle parameters</li> <li>• BI (unless parametric release)</li> </ul>
Volumes	High	Limited	High
Turn time	Fast (<24 hours)	Very Fast (<8 hours)	Long (1 week)

You should now feel better at ease with the following concepts :

- What is sterilization Vs decontamination
- Aseptic Assembly Vs Terminal sterilization
- Irradiation sterilization (Gamma and E-beam)
- Ethylene Oxide sterilization

- *ISO 11135:2014* Sterilization of medical devices – Requirements for the development; validation and routine Control of a Sterilization Process for Medical Devices – Ethylene Oxide
- *ISO 10993-7:2008 (R) 2012* Biological evaluation of medical devices - Part 7: Ethylene oxide sterilization residuals
- *ISO 11137-1* Sterilization of health care products – Radiation – Part 1: Requirements for development, validation, and routine control of a sterilization process for medical devices
- *ISO 11137-2* Sterilization of health care products – Radiation – Part 2: Establishing the sterilization dose
- *ISO 11737-1:2018* Sterilization of medical devices (Microbiological methods) Part 1: Determination of a population of microorganisms on products
- *ISO 11737-2:2009 (R) 2014*
- Sterilization of medical devices (Microbiological methods) Part 2: Tests of sterility performed in the definition, validation and maintenance of a sterilization process
- *ISO 11138-1:2017*
- Sterilization of health care products (Biological indicators) Part 1: General requirements
- *ISO 11138-2:2017*
- Sterilization of health care products (Biological indicators)Part 2: Biological indicators for ethylene oxide sterilization processes
- *ISO 14161: 2009 (R) 2014*
- Biological indicators. Guidance for the selection, use and interpretation of results

- *ISO 11737-2:2009 (R) 2014*  
Sterilization of medical devices (Microbiological methods) Part 2: Tests of sterility performed in the definition, validation and maintenance of a sterilization process
- *ISO TS 19930:2017 Guidance on aspects of a risk-based approach to assuring sterility of a terminally-sterilized, single use health care product unable to withstand processing to achieve maximally a sterility assurance level of 10<sup>-6</sup>*
- *AAMI TIR 33 Sterilization of health care products—Radiation—Substantiation of a selected sterilization dose — Method Vdmax*
- *United States Pharmacopeia (USP) Chapter <71> Sterility Tests*
- *Eudralex Volume 4 – GMP Annex 1*
- *Eudralex Volume 4 – GMP Annex 12*
- *European Pharmacopeia (EP) Chapter 2.6.1 Sterility*
- *The Aseptic and Sterile Processing: Control, Compliance and Future Trends* - Edited by Tim Sandle, Edward Tidswell PDA – 2017
- *PDA Survey: 2017 PDA Aseptic Processing*
- *A comparison of Gamma, E-beam, X-Ray and ETO technologies for the industrial Sterilization of MD and Health care products – GIPA, IIA – 31 Aug 2017*



[agillet@eu.sterigenics.com](mailto:agillet@eu.sterigenics.com)