

# The Syringe

*Example Specification – syringe body, RTF*

Bernd Zeiss

*Venice, April 20<sup>st</sup> and 21<sup>st</sup> 2023*



## Agenda – DAY 2

### **The Syringe (Body)**

*Example Specification*

### **Plunger stoppers, Needle Shields, Tip caps**

*Materials and Properties, Functionality, production; Extractables*

### **Manufacturing Aspects in Fill & Finish and Assembly**

*Bulk versus Nested, Nest Sizes, Rod insertion, Handling of Syringes, Labeling, glass to glass contact*

### **Hands-on Session 2, Mind map, Lottery**

# Requirements

## Primary Packaging Containers

*Recap from yesterday*

### Consider container materials from the beginning

- Critical contact material
- Device (functions) and container at the same time

### Fill and Finish compatibility

- Standardization needed
- Transparency (visual inspection)
- Sterilization



# Requirements Primary Packaging Containers

*Recap from yesterday*

## Physical characteristics

- Material/ breakability
- Design, size, wall thickness
- standardized by ISO
  
- Long term storage
- Opening forces
- Delivery forces
- Closure integrity (CCI)

## Pharmaco-chemical characteristics

- USP, EP, JP tests
- L&E, Trace metals, impurities
  
- Stability/interactions with drug substance
- Endotoxin level
- Biocompatibility
- Subvisible particles




# Requirements Primary Packaging Containers

*Recap from yesterday*

## Patient /end user

- Functionality
  - Cap removal, gliding force, needle penetration pain
- Administration volume
- Autoinjectors
- Safety syringes and needle stick prevention
- Frequency of administration
  
- Intended use
- Market
- Pricing




# Syringe Example Specification 1

Contains and describes all relevant RTF syringe information

Agreed on and signed between supplier and pharma company

- Product
- Product Packaging
- Shipment Packaging
- Labeling
- Storage
- Shelf Life
- Quality System
- Regulatory Compliance
- Sampling
- Documentation
- Attachments
- Customer Responsibilities



<b>Volume</b>	1.0 mL long
<b>Cannula/ Cone</b>	staked needle, 27G, 1/2 inch, 3B NW
<b>Closure</b>	RNS grey Stelmi 4800GS
<b>Finger Flange</b>	Cut Finger Flange
<b>Siliconization</b>	Oily siliconized: 0.5 mg
<b>Nest size</b>	160 syringes
<b>Packaging – Bag</b>	Single Bag
<b>Packaging – Pallet</b>	Euro Pallet

What is missing?

- no plunger stopper, no rod, no drug
- RTF= „ready to fill“, i.e. *one component* of the final filled and stoppered syringe
- empty

# Syringe Example Specification 2

## Materials (glass, polymer)


The syringe is made from

- borosilicate glass, hydrolytic resistance Type 1
- COP or COC

## Shape, Diameter of syringe

- Properties and dimensions are in accordance with DIN ISO 11040, if nothing different is specified
- Dimensions are according to the drawing no. xxx

0.5 ml	47.6	6.85	4.65	1/2
1.0 ml long	54.0	8.15	6.35	1/2
1.0 ml standard	35.7	10.85	8.65	1/2   5/8
1.5 ml	43.2	10.85	8.65	1/2   5/8
2.25 ml	54.4	10.85	8.65	1/2






# Syringe Example Specification 3

## Cone

- Integrated needle syringe, luer lock adapter or luer cone - all with cap
- Description of cannula e.g. 27G 1/2"



## Needle shield, tip cap

- Elastomer cap described in specification - rubber formulation, design



Free Needle Length	Outer Diameter	Inner Diameter	Grinding
5/8" [15.9 mm]	23 G [0.64 mm]	[0.41 mm] thin walled	3 Bevel
5/8" [15.9 mm]	25 G [0.5 mm]	[0.25 mm] standard [0.32 mm] thin walled	3 Bevel
5/8" [15.9 mm]	26 G [0.45mm]	[0.24mm]	3 Bevel
1/2" [12.7 mm]	27 G [0.41 mm]	[0.21 mm] standard	3 Bevel
	27 G [0.425 mm]	[0.20 mm] standard	3 Bevel XL
	27 G [0.40 mm]	[0.24 mm] standard [0.27 mm] thin wall [0.28 mm] thin wall	V®-bevel
1/2" [12.7 mm]	29 G [0.33 mm]	[0.19 mm] thin wall	3 Bevel
		[0.20 mm] standard	3 Bevel XL V®-bevel



# Syringe Example Specification 4

## Mechanical properties



- Special mechanical tests can be defined:
  - cone breakage, shoulder breakage, finger flange breakage
- Needle shield or cap pull-off force range given

## Siliconization

- Often Dow Corning DC 360
- Special siliconization types defined e.g. baked-on siliconization (to meet USP 789)
- Needle siliconization, glue description

## Chemical Properties

- Borosilicate glass Type 1, supplier name often stated
- Tungsten levels can be specified
- EtO - treatment according to DIN EN ISO 11135 and ISO 10993-7- Ethylene Oxide Sterilization Residuals



# Syringe Example Specification 5

**Quality and Regulatory guidelines and technical standards: EU / US / ISO**

ISO 9001 “Quality Management Systems - Requirements”

ISO 15378 “Primary packaging materials for medicinal products - Particular requirements for the application of ISO 9001:2008, with reference to Good Manufacturing Practice (GMP)”

21 CFR 211, Subpart E “Current Good Manufacturing Practice for Finished Pharmaceuticals”

21 CFR 820 „Quality System Regulation - Medical Devices“

ISO 13485 „Medical Devices - Quality Management Systems“



# Syringe Example Specification 6

## Sterility

- Ph. Eur. 2.6.1 USP <71>
- Sterility Tests according to ICH Q4B Annex 8

## Bacterial Endotoxins

- Ph. Eur. 2.6.14 and USP <85>

## Glass container

- Ph. Eur. 3.2.1, USP <660>, JP 7.01
- ISO 4802-2 Hydrolytic Resistance, Container Class HC1
- ISO 11040-4 “Prefilled syringes - Part 4: Glass barrels for injectables”
- Ph. Eur. 2.4.20, USP <211> Arsenic

## Cannula

- ISO 9626 “Stainless steel needle tubing for the manufacture of medical devices”



# Syringe Example Specification 7

## RNS and tip caps: Rubber formulation

- Ph. Eur. 3.2.9, USP <381>
- ISO 8871 “Elastomeric parts for aqueous parenteral preparations”
- USP <87>, USP <88>: “Biological Reactivity Tests”

## Lubricant


- Conformity to applicable Monographs of EP and USP

## Adhesive

- USP <88> “Biological Reactivity Tests, in Vivo”

## General

- BSE/TSE
- CONEG Toxic Packaging legislation
- Mercury and hexavalent Chromium



# Syringe Example Specification 8

## Drawings and description of packaging

- Syringe
- Nest, tub
- outer packaging and labeling
- pallet

## Approval page

- Signatures of supplier
- Signatures of customer



## Storage and shelf life information

## Sampling

- According to DIN ISO 2859-1

## Documentation

- Certificate of Conformance (CoC) with each delivery
- The CoC states the batch information
  - conformance of all components
  - EtO residues
  - Sterility
  - Endotoxin level of the syringe



# Syringe Example Specification 9

## Defect evaluation lists

- describe possible defects

## Acceptance Quality Limit – AQL

- Define Pharma quality criteria
- “Quality level that is the worst tolerable” ISO 2859-1 - Give a probability of defects

## Possible defects:

- Packaging e. g. labeling or damaged tub
- Syringe barrel e.g. cracks or deformations
- Dimensional defects e.g. of total length with cap
- Cannula e.g. hooks or contamination
- Closure e.g. pull-off force oot or pierced
- Chemical tests esp. hydrolytic resistance
- Sterility issues after EtO
- Siliconization e.g. glide force oot

**How many samples** should be picked and inspected among a batch of product or parts?

**Where is the limit** between acceptability and refusal when it comes to defective products?

Example:

“I want no more than 1.5% defective items in the whole order quantity, on average over several production runs with that supplier” means the AQL is 1.5%.

For most consumer goods, the limits are:



- 0% for critical defects (totally unacceptable: a user might get harmed, or regulations are not respected).
- 2.5% for major defects (these products would usually not be considered acceptable by the end user).
- 4.0% for minor defects (there is some departure from specifications, but most users would not mind it).

This tool is used during final inspections (before the products are ready to be shipped out)

# Possible Defects



**Technical Report No. 43 (Revised 2013)**  
Identification and Classification of  
Nonconformities in Molded and Tubular Glass  
Containers for Pharmaceutical Manufacturing:  
Covering Ampoules, Bottles, Cartridges,  
Syringes and Vials






# Syringe Example Specification

**Back stops**  
**Rods**  
**Plunger stoppers**

- Separate specifications
- Dimensions
- Material
- Pack size



# Test procedures

[11040-4 ISO](#)  
[11040-7 ISO](#)  
[11040-8 ISO](#)  
[80369-7 ISO](#)



# CoC and statements

## Completing the syringe specification

- CoC Certificate of Conformance with every delivered batch
- Statements regarding Elemental Impurities, TSE, BSE, REACH, debarment...on demand




# Limits of a Syringe Specification

## Drug properties

### Not part of the syringe specification

- Impact of different drug properties
- Stability testing with drug
- Risk of drug interactions with syringe components
- Risk of impaired functionality
- Leacheables test
- Syringe is not a combination product - yet



# Limits of a Syringe Specification

## Drug properties

Prefillable syringe becomes a combination product after filling

MDR not fully applicable

Usability studies not fully covered with empty syringes


### Comprehensive tests need to be carried out at the Pharma company

- Stability testing with drug – see e.g. [PDA report 73](#)
- Mechanical tests e.g. with Autoinjector and drug
- Performance with formulation (e.g. glide force, break tests), particles to be tested with drug formulation



**Technical Report No. 73**  
Prefilled Syringe User Requirements for  
Biotechnology Applications





# Summary – Syringe specification

- Specification is the „contract“ between supplier and pharma company
- Highly standardized: all syringe barrel features included
- Prefillable (primary packaging) features covered
- Not covering syringe system functionality (filled)
- Certificate of Conformance for every batch delivered
- Closer cooperation necessary in future to cover system functionality